

VERIFICATION OF CALIBRATION REPORT

Of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory – Statewide Breath Alcohol Program

DataMaster cdm S/N 130143

Supervisor/Operator Performing the Verification Procedures:

Name: Morgan Boatman ID# 7386 Date 26 Aug 2011

Agency Directorate of Emergency Services, Fort Wainwright, Alaska Phone# 353-9011

Instrument Location Bldg #3028, Fort Wainwright, Alaska

Alco S/N X301540 Target Value .083 High Pressure _____

Alco Test Value Average 0.088 ^{BMB} 9/2/11 0.087 ^{BMB} 9/2/11
2nd Alco

Signature [Signature]

(OVER)

BMB
9/2/11

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 2 day of Sept, 2010. [Signature]

[Signature]
Nikki Roth, Notary Public
My Commission Expires With Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130143

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130143

AUGUST 26, 2011

OPERATOR'S NAME:
BOATMAN

OPERATOR'S NUMBER: 7386

SUBJECT'S LAST NAME:

VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :

L

O.L. #: L

DEPT/AGENCY: FA14

CASE/REPORT:

TEST TYPE: U

ALCO TARGET VALUE: .003

ALCO S/N: 130143

— BREATH ANALYSIS —

.003 ADJUSTED FOR 29.34 in

ALCO TARGET .001 16:47

BLANK TEST .000 16:48

INTERNAL STANDARD VERIFIED 16:48

ALCO TV 29.34 in .003 16:49

BLANK TEST .000 16:50

SUBJECT SAMPLE .000 16:50

BLANK TEST .000 16:51

ALCO TV 29.34 in .007 16:52

BLANK TEST .000 16:52

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130143

AUGUST 26, 2011

TIME 16:31

— DIAGNOSTIC CHECK —

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/28/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 43c

BAROMETER: 29.34 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@ABCDEF
GHIJKL*NOPQRSTU VWXYZ[]^_`ab cdefghijklmno
pqrstuvwxyz{|}~